

Memorandum

FROM : Director, Utilities Division, Facilities Maintenance Department,
MCAS, Cherry Point, NC 28533

TO : Director, Natural Resource and Environmental Affairs Division, Base
Maintenance, MCB, Camp Lejeune, NC 29542

SUBJ : Pure Water Samples; analysis of

Ref : (a) Safe Drinking Water Act (SDWA), Public Law #93-523

Encl : (1) Sample Analysis Report

1. The pure water samples from your laboratory have been analyzed in accordance with laboratory certification requirements by the reference, and the results are provided in the enclosure.

2. It will be noted that certain limitations have been exceeded as follows:

<u>Sample No.</u>	<u>Parameter</u>	<u>Results</u>	<u>Limit</u>
2	Copper	0.27 mg/l	0.05 mg/l
1	Zinc	0.10 mg/l	0.05 mg/l
1	Nickel	0.11 mg/l	0.05 mg/l

3. It is recommended that the distillation units be cleaned and the pure water be retested.

RE Hall
R. E. HALL
Acting

CLW

0000000417